

-80°C Freezers, Backup

PSC 543, 555, 559, 637, NSC 488, 340

Contact: [Sandy Hsieh 404-413-5363; yhsieh@gsu.edu](mailto:yhsieh@gsu.edu)

1. Backup freezers can only be used for temporary storage.
 - a. Lab -80oC freezer is broken. Store until repaired.
 - b. Defrost Lab -80oC freezer. Store until Lab freezer back up to temperature.
2. Put a sign on door of freezer with your name, Lab contact information, reason for storage and an approximate time you will be out of the freezer.
3. Contact core facility personnel to co-sign on the new sign.