UV Crosslinker (PSC 559)

Every time you use the UV Crosslinker, you MUST sign the log book.

Contact: Sonja Young (404) 413-5363; sstovall@gsu.edu

- 1. Power on
- 2. place sample inside (on a clean kimwipe)
- 3. Optimal crosslink
- 4. Start
- 5. Reset
- 6. Start